

**MINISTÉRIO DAS OBRAS PÚBLICAS, TRANSPORTES E COMUNICAÇÕES**  
**Laboratório Nacional de Engenharia Civil**

Departamento de Materiais

Proc. 0204/11/16162

Proc. 0204/75/8104

---

**QUALANOD INTER-LABORATORY TESTS OF ANODIZED ALUMINIUM TESTING METHODS**  
**STATISTICAL ANALYSIS - DETERMINATION OF REPEATABILITY AND REPRODUCIBILITY**  
**REPORT 297/2006 - DM**

Lisboa, Outubro 2006

---


Trabalho realizado para

QUALANOD (EURAS/EWAA)



# QUALANOD INTER-LABORATORY TEST OF ANODIZED ALUMINIUM TESTING METHODS

## *STATISTICAL ANALYSIS - REPEATIBILITY AND REPRODUCIBILITY DETERMINATION*





## Institutions/laboratories that intended to participate in this inter-laboratory test

Country	Name	Laboratory
AUSTRIA	Aluminium Ranshofen Service GmbH	AMAG
BELGIUM	CORI - Coatings Research Institute	CoRI
FRANCE	TESTAL	-
GERMANY	IFO GmbH - Institut für Oberflächentechnik	IO GmbH
GREECE	AAG Quality – EKANAL for Aluminium Ass. Of Greece	EKANAL
HUNGARY	ÉMI KHT - Institut für Qualitätskontrolle	ÉMI KHT
ITALY	QUALITAL	QUALITAL
NETHERLANDS	Adviescentrum VOM B.H.	A VOM B. V.
POLAND	Instytut Mechaniki Precyzyjnej	IMP
PORTUGAL	LNEC - Laboratório Nacional de Engenharia Civil	LNEC
SPAIN	QUALESPAÑA - Ministerio de Vivienda Subdirección General de Innovación y Calidad de la Edificación	CEDEX
SWITZERLAND	EMPA, Abtl. Korrosion	EMPA
TURKEY	Turkish Standards Institution	-

Laboratories from Turkey and France haven't sent any results

## Testing methods

- **EN ISO 2360:2003** - *Non-conductive coatings on non-magnetic electrically conductive basis materials. **Measurement of coating thickness.** Amplitude-sensitive eddy current method (ISO 2360:2003)*
- **EN 12373-7:2002** (2<sup>nd</sup>Ed.) - *Aluminium and aluminium alloys. Anodizing. Part 7: Assessment of **quality of sealed anodic oxidation coatings by measurement of the loss of mass** after immersion in phosphoric acid/chromic acid solution with prior acid treatment*
- **EN 12373-5:1998** - *Aluminium and aluminium alloys. Anodizing. Part 5: Assessment of **quality of sealed anodic oxidation coatings by measurement of admittance***
- **EN 12373-4:1998** - *Aluminium and aluminium alloys. Anodizing. Part 4: Estimation of loss absorptive power of anodic oxidation coatings after sealing by **dye spot test** with prior acid treatment*



## General characteristics of the test specimens

ALUMINIUM	Alloy Type	Type of product	Dimensions of testing specimens	
	6060	Sheet with 1 mm of thickness	70mmx140 mm	
ANODIC COATING	Type	Thickness class	Sealing	Colour
Polished surface	<b>P</b>	15 $\mu\text{m}$	3 min/ $\mu\text{m}$	natural
	<b>IP</b>	15 $\mu\text{m}$	not sealed	natural
	<b>I</b>	25 $\mu\text{m}$	1 min/ $\mu\text{m}$	bronze
Satined surface	<b>A</b>	20 $\mu\text{m}$	3 min/ $\mu\text{m}$	natural

Fifty specimens of each coating type were produced




## Anodizing and sealing conditions used for the production of test specimens

Coating type	Anodizing bath		Sealing bath	
P	Free H <sub>2</sub> SO <sub>4</sub> -	184,7 g/l	Demineralised water	
	Al content -	12,8 g/l	pH -	5,70
	Temperature -	18,0 °C	Additive -	Anodal SH1
	Current density -	1,6 A/dm <sup>2</sup>	Time -	3 min/μm
I	Free H <sub>2</sub> SO <sub>4</sub> -	182,1 g/l	Demineralised water	
	Al content -	11,4 g/l	pH -	5,75
	Temperature -	19,0 °C	Additive -	Anodal SH1
	Current density -	1,7 A/dm <sup>2</sup>	Time -	1 min/μm
IP	Free H <sub>2</sub> SO <sub>4</sub> -	184,7 g/l	Not sealed	
	Al content -	12,8 g/l		
	Temperature -	19,0 °C		
	Current density -	1,6 A/dm <sup>2</sup>		
A	Free H <sub>2</sub> SO <sub>4</sub> -	184,7 g/l	Demineralised water	
	Al content -	12,8 g/l	pH -	5,70
	Temperature -	19,0 °C	Additive -	Anodal SH1
	Current density -	1,7 A/dm <sup>2</sup>	Time -	3 min/μm

# Instructions

- Laboratory code number
- Measuring points




Scheme of the testing zones for thickness measurement



Sample with 70mmX50mm for sealing assessment by the mass loss test

Scheme of testing zones for sealing quality assessment





# EXCEL Worksheets for data registration

## Protocol

Test laboratory:	LNEC
Sample type:	P
Tested by:	N Garcia
Date of report:	

## Test conditions

Temperature:	21	± 2	°C
Date of the test:	05-12-2005		yyyy-mm-dd
Zero base	0,0	±	µm
Calibration standard 1	23,5	± 0,5	µm
Calibration standard 2		±	µm
Calibration standard 3		±	µm

## Test results

### Thickness results (ISO 2360)

Sample No.	Front side (µm)			
	1st point	2nd point	3rd point	
P15				
measure 1	21,6			
measure 2	22,0			
measure 3	21,3			
Average thickness	21,6			21,6

### Weight loss results (EN 12373-7)

Sample No.	Height H	Lenght L	Sample thickness T	Area A	Weigth loss
	(mm)	(mm)	(mm)	(dm <sup>2</sup> )	(mg/dm <sup>2</sup> )
P15					
measure 1	71,59	50,29	1,04	-	-
measure 2	71,34	50,86	1,05	-	-
measure 3	71,83	51,08	1,03	-	-
Average	71,59	50,74	1,04	0,7506	15

### Admittance results (EN 12373-5)

Sample No.	Measurement	Thickness	Temperature	Admittance
	Ym	(µm)	(°C)	Y
P15				
	10,5	24,6	21,2	8,0

### Dye spot results (EN 12373-4)

Dye spot	0-1
----------	-----

Sample No.	Front side (µm)			
	1st point	2nd point	3rd point	
measure 1				
measure 2				
measure 3				
Average thickness				

Sample No.	Height H	Lenght L	Sample thickness T	Area A	Weigth loss
	(mm)	(mm)	(mm)	(dm <sup>2</sup> )	(mg/dm <sup>2</sup> )
P15					
measure 1				-	-
measure 2				-	-
measure 3				-	-
Average					

Sample No.	Measurement	Thickness	Temperature	Admittance
	Ym	(µm)	(°C)	Y

Dye spot	
----------	--

Sample No.	Front side (µm)			
	1st point	2nd point	3rd point	
measure 1				
measure 2				
measure 3				
Average thickness				

Sample No.	Height H	Lenght L	Sample thickness T	Area A	Weigth loss
	(mm)	(mm)	(mm)	(dm <sup>2</sup> )	(mg/dm <sup>2</sup> )
measure 1				-	-
measure 2				-	-
measure 3				-	-
Average					

Sample No.	Measurement	Thickness	Temperature	Admittance
	Ym	(µm)	(°C)	Y

Dye spot	
----------	--


Example for the registration of tests results for specimens of coating type P




# Results reported



## Test results of method EN ISO 2360




## Test results of method EN 12373-7




# Results reported



## Test results of method EN 12373-5



## Test results of method EN 12373-4





## Precision analysis according to ISO 5725-2

- Consistency tests
  - Graphical consistency technique – Mandel's  $h$  and  $k$  statistics

$$h_{ij} = \frac{\bar{y}_{ij} - \bar{y}_j}{\sqrt{\frac{1}{(p_j - 1)} \sum (\bar{y}_{ij} - \bar{y}_j)^2}} \quad k_{ij} = \frac{s_{ij} \sqrt{p_j}}{\sqrt{\sum s_{ij}^2}}$$

- Numerical outlier technique – Cochran's and Grubb's tests

$$C = \frac{s_{max}^2}{\sum_{i=1}^p s_i^2}$$

$$G_p = (x_p - \bar{x})/s$$

$$G_1 = (\bar{x} - x_1)/s$$

$$G = s_{p-1,p}^2 / s_0^2$$

$$G = s_{1,2}^2 / s_0^2$$

- Calculation of the **general mean and variances**

$$\hat{m}_j = \bar{y}_j = \frac{\sum_{i=1}^p n_{ij} \bar{y}_{ij}}{\sum_{i=1}^p n_{ij}}$$

$$s_{ij}^2 = \frac{\sum_{i=1}^p (n_{ij} - 1) s_{ij}^2}{\sum_{i=1}^p (n_{ij} - 1)}$$


*Repeatability variance*

$$s_{Rj}^2 = s_{ij}^2 + s_{Lj}^2$$

*Reproducibility variance*

$$\left[ s_{dj}^2 = \frac{1}{p-1} \sum_{i=1}^p n_{ij} (\bar{y}_{ij} - \bar{y}_j)^2 = \frac{1}{p-1} \left[ \sum_{i=1}^p n_{ij} (\bar{y}_{ij})^2 - (\bar{y}_j)^2 \sum_{i=1}^p n_{ij} \right] \right]$$

# Thickness measurement (EN ISO 2360)



# Thickness measurement (EN ISO 2360)



## Laboratories outside critical value lines of Mandel's statistics

Level	P	IP	I	A
Mandel's <i>k</i> -plot	Lab 10	-	-	Lab 11; Lab 12
Classification	<i>Straggler</i>	-	-	<i>Straggler</i>
Mandel's <i>h</i> -plot	Lab 11	-	-	-
Classification	<i>Outlier</i>	-	-	-

## Cochran's test results

Level	P	IP	I	A
Valid laboratories <i>p</i>	11	11	11	11
Number of replicates <i>n</i>	3	3	3	6
1% Critical value $C_{Cr} (1\%)$	0,504	0,504	0,504	0,332
5% Critical value $C_{Cr} (5\%)$	0,417	0,417	0,417	0,281
Cochran's test statistic <b>C</b>	0,266	0,148	0,222	0,206
Classification	<i>correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $C > C_{Cr} (1\%)$ )	-	-	-	-



# Thickness measurement (EN ISO 2360)

## Grubb's test results

Level	P	IP	I	A
Valid laboratories $p$	11	11	11	11
Single $G_{Cr}$ (1%)	2,564	2,564	2,564	2,564
Single $G_{Cr}$ (5%)	2,355	2,355	2,355	2,355
Single high $G_p$	0,926	1,650	1,658	1,512
Single low $G_1$	2,328	1,627	1,308	1,430
Classification (low)	<i>correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_p > G_{Cr}$ (1%))	-	-	-	-
Classification (low)	<i>correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_1 > G_{Cr}$ (1%))	-	-	-	-
Double $G_{Cr}$ (1%)	0,1448	0,1448	0,1448	0,1448
Double $G_{Cr}$ (5%)	0,2213	0,2213	0,2213	0,2213
Double high $G_{largest}$	0,826	0,5152	0,4244	0,5539
Double low $G_{smallest}$	0,1084	0,3662	0,5964	0,5738
Classification (two largest)	<i>correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_{largest} < G_{Cr}$ (1%))	-	-	-	-
Classification (two smallest)	<b>Outliers</b>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_{smallest} < G_{Cr}$ (1%))	<b>Lab 3; Lab 11</b>	-	-	-

Single: test for one outlying observation; Double: test for two outlying observations

# Thickness measurement (EN ISO 2360)





## Results of precision analysis

Level	P	IP	I	A
Number of replicates $n$	3	3	3	6
Valid laboratories $p$	11	11	11	11
<b>General mean <math>m / \mu\text{m}</math></b>	<b>18,64</b>	<b>19,49</b>	<b>27,78</b>	<b>21,35</b>
Repeatability variance $s_r^2$	0,333	0,321	0,312	0,294
Between-lab variance $s_L^2$	0,511	0,928	1,966	0,493
Reproducibility variance $s_R^2$	0,844	1,248	2,278	0,788
<b>Repeatability std. dev. <math>s_r</math></b>	<b>0,58</b>	<b>0,57</b>	<b>0,56</b>	<b>0,54</b>
<b>Reproducibility std. dev. <math>s_R</math></b>	<b>0,92</b>	<b>1,12</b>	<b>1,51</b>	<b>0,89</b>
Repeatability COV ( $s_r/m$ )	3,1%	2,9%	2,0%	2,5%
Reproducibility COV ( $s_R/m$ )	4,9%	5,7%	5,4%	4,2%
Number of outliers	2	0	0	0
Number of excluded outliers	0	0	0	0
Outlier type	<sup>1</sup> Mh, <sup>2</sup> G(II)	-	-	-
Outlier laboratories	Lab 3 <sup>2</sup> Lab 11 <sup>1,2</sup>	-	-	-

Outlier type: Mh – Mandel's  $h$ ; Mk – Mandel's  $k$ ; C - Cochran's; G(I) – Grubs (one outlying observation); G(II) – Grubs (two outlying observations)

# Sealing quality assessment by mass loss (EN 12373-7)





# Sealing quality assessment by mass loss (EN 12373-7)



## Laboratories outside critical value lines of **Mandel's** statistics

Level	P	IP	I	A
<i>Mandel's k-plot</i>	Lab 10	Lab 9	Lab 8	Lab 8
Classification	<i>Outlier</i>	<i>Outlier</i>	<i>Outlier</i>	<i>Outlier</i>
<i>Mandel's h-plot</i>	Lab 7	Lab 4	Lab 7	Lab 7
Classification	<i>Outlier</i>	<i>Straggler</i>	<i>Straggler</i>	<i>Outlier</i>

## **Cochran's** test results

Level	P	IP	I	A
Valid laboratories $p$	11	11	11	11
Number of replicates $n$	3	3	3	3
1% Critical value $C_{Cr(1\%)}$	0,504	0,504	0,504	0,504
5% Critical value $C_{Cr(5\%)}$	0,417	0,417	0,417	0,417
Cochran's test statistic $C$	0,731	0,400	0,583	0,515
Classification	<i>Outlier</i>	<i>correct</i>	<i>Outlier</i>	<i>Outlier</i>
Outlier Lab ( $C > C_{Cr(1\%)}$ )	Lab 10	-	Lab 8	Lab 8

# Sealing quality assessment by mass loss (EN 12373-7)



## Grubb's test results

Level	P	IP	I	A
Valid laboratories $p$	11	11	11	11
Single $G_{Cr}$ (1%)	2,564	2,564	2,564	2,564
Single $G_{Cr}$ (5%)	2,355	2,355	2,355	2,355
Single high $G_p$	2,320	1,181	2,213	2,417
Single low $G_1$	1,746	2,040	1,512	1,299
Classification (high)	<i>Correct</i>	<i>correct</i>	<i>correct</i>	<i>Straggler</i>
Straggler Lab ( $G_p > G_{Cr}$ (5%))	-	-	-	Lab 7
Classification (low)	<i>Correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_1 > G_{Cr}$ (1%))	-	-	-	-
Double $G_{Cr}$ (1%)	0,1448	0,1448	0,1448	0,1448
Double $G_{Cr}$ (5%)	0,2213	0,2213	0,2213	0,2213
Double high $G_{largest}$	0,3530	0,7224	0,3164	-
Double low $G_{smallest}$	0,5480	0,2748	0,5986	0,6591
Classification (two largest)	<i>Correct</i>	<i>correct</i>	<i>correct</i>	-
Outlier Lab ( $G_{largest} < G_{Cr}$ (1%))	-	-	-	-
Classification (two smallest)	<i>Correct</i>	<i>correct</i>	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_{smallest} < G_{Cr}$ (1%))	-	-	-	-

Single: test for one outlying observation; Double: test for two outlying observations

# Sealing quality assessment by mass loss (EN 12373-7)




## Results of precision analysis

Level	P	IP	I	A
Number of replicates $n$	3	3	3	3
Valid laboratories $p$	10	11	10	10
<b>General mean <math>m</math> / g.dm<sup>-2</sup></b>	<b>15,92</b>	<b>477,5</b>	<b>17,86</b>	<b>13,19</b>
Repeatability variance $s_r^2$	0,426	140,55	0,457	0,430
Between-lab variance $s_L^2$	8,359	141,11	9,708	7,442
Reproducibility variance $s_R^2$	8,785	281,66	10,164	7,871
<b>Repeatability std. dev. <math>s_r</math></b>	<b>0,65</b>	<b>11,7</b>	<b>0,68</b>	<b>0,66</b>
<b>Reproducibility std. dev. <math>s_R</math></b>	<b>2,96</b>	<b>16,8</b>	<b>3,19</b>	<b>2,81</b>
Repeatability COV ( $s_r/m$ )	4,1%	2,5%	3,8%	5,0%
Reproducibility COV ( $s_R/m$ )	18,6%	3,5%	17,9%	21,3%
Number of outliers	2	1	1	3
Number of excluded outliers	1	0	1	1
Outlier type	<sup>1</sup> Mh, <sup>2</sup> Mk, <sup>3</sup> C	Mk	Mk, C	<sup>1</sup> Mh, <sup>2</sup> Mk, <sup>3</sup> C
Outlier laboratories	Lab 7 <sup>1</sup> Lab 10 <sup>1,2,3</sup>	Lab 9	Lab 8	Lab 7 <sup>1</sup> Lab 8 <sup>2,3</sup>

Outlier type: Mh – Mandel's  $h$ ; Mk – Mandel's  $k$ ; C - Cochran's; G(I) – Grubs (one outlying observation); G(II) – Grubs (two outlying observations)

# Sealing quality assessment by measurement of admittance (EN 12373-5)





# Sealing quality assessment by measurement of admittance (EN 12373-5)

## Laboratories outside critical value lines of **Mandel's** statistics

Level	P	IP	I	A
<i>Mandel's k-plot</i>	Lab 12	n.a.	Lab 6	Lab 8
Classification	<i>Outlier</i>	n.a.	<i>Outlier</i>	<i>Outlier</i>
<i>Mandel's h-plot</i>	Lab 8	n.a.	Lab 8	Lab 4
Classification	<i>Straggler</i>	n.a.	<i>Straggler</i>	<i>Straggler</i>

n. a. – not applied

## **Cochran's** test results

Level	P	IP	I	A
Valid laboratories $p$	11	n.a.	11	11
Number of replicates $n$	3	n.a.	3	3
1% Critical value $C_{Cr (1\%)}$	0,504	n.a.	0,504	0,504
5% Critical value $C_{Cr (5\%)}$	0,417	n.a.	0,417	0,417
Cochran's test statistic $C$	0,893	n.a.	0,419	0,862
Classification	<i>Outlier</i>	n.a.	<i>Straggler</i>	<i>Outlier</i>
Outlier Lab ( $C > C_{Cr (1\%)}$ )	Lab 12	n.a.	Lab 6	Lab 8

n. a. – not applied



# Sealing quality assessment by measurement of admittance (EN 12373-5)

## Grubb's test results

Level	P	IP	I	A
Valid laboratories $p$	11	n.a.	11	11
Single $G_{Cr}$ (1%)	2,564	n.a.	2,564	2,564
Single $G_{Cr}$ (5%)	2,355	n.a.	2,355	2,355
Single high $G_p$	1,352	n.a.	1,497	1,961
Single low $G_1$	2,099	n.a.	1,884	1,642
Classification (high)	<i>correct</i>	n.a.	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_p > G_{Cr}$ (1%))	-	n.a.	-	-
Classification (low)	<i>correct</i>	n.a.	<i>correct</i>	<i>correct</i>
Outlier Lab ( $G_1 > G_{Cr}$ (1%))	-	n.a.	-	-
Double $G_{Cr}$ (1%)	0,1448	n.a.	0,1448	0,1448
Double $G_{Cr}$ (5%)	0,2213	n.a.	0,2213	0,2213
Double high $G_{largest}$	0,6505	n.a.	0,4987	0,4826
Double low $G_{smallest}$	0,3741	n.a.	0,4238	0,5148
Classification (two largest)	<i>correct</i>	n.a.	<i>correct</i>	<i>correct</i>
Outlier Lab( $G_{largest} < G_{Cr}$ (1%))	-	n.a.	-	-
Classification (two smallest)	<i>correct</i>	n.a.	<i>correct</i>	<i>correct</i>
Outlier Lab( $G_{smallest} < G_{Cr}$ (1%))	-	n.a.	-	-

Single: test for one outlying observation; Double: test for two outlying observations  
 n. a. – not applied

# Sealing quality assessment by measurement of admittance (EN 12373-5)




## Results of precision analysis

Level	P	I	A
Number of replicates $n$	3	3	3
Valid laboratories $p$	10	11	10
<b>General mean <math>m / Y</math></b>	<b>8,98</b>	<b>15,29</b>	<b>9,88</b>
Repeatability variance $s_r^2$	0,113	0,747	0,113
Between-lab variance $s_L^2$	2,858	8,974	1,794
Reproducibility variance $s_R^2$	2,970	9,720	1,907
<b>Repeatability std. dev. <math>s_r</math></b>	<b>0,34</b>	<b>0,86</b>	<b>0,34</b>
<b>Reproducibility std. dev. <math>s_R</math></b>	<b>1,72</b>	<b>3,12</b>	<b>1,38</b>
Repeatability COV ( $s_r/m$ )	3,7%	5,7%	3,4%
Reproducibility COV ( $s_R/m$ )	19,2%	20,4%	14,0%
Number of outliers	2	1	1
Number of excluded outliers	1	0	1
Outlier type	<sup>1</sup> Mh, <sup>2</sup> Mk, <sup>3</sup> C	Mk	Mk, C
Outlier laboratories	Lab 8 <sup>1</sup> Lab 12 <sup>2,3</sup>	Lab 6	Lab 8

Outlier type: Mh – Mandel's  $h$  ;Mk – Mandel's  $k$  ; C - Cochran's ; G(I) – Grubs (one outlying observation) ; G(II) – Grubs (two outlying observations)

# Dependency analysis of precision (repeatability and reproducibility) with the mean





# Precision analysis results



EN ISO 2360 – Thickness measurement	Coating type				Excluded data lab:coating type
	P	IP	I	A	
General mean $m / \mu\text{m}$	18,6	19,5	27,8	21,4	None
Repeatability std. dev. $s_r$	0,58	0,57	0,56	0,54	
Reproducibility std. dev. $s_R$	0,92	1,12	1,51	0,89	
Global repeatability std. dev.	0,56				
Global reproducibility std. dev.	$S_R = 0,051 m (R^2=0,7338)$				
EN 12373-7 - Sealing quality assessment by mass loss	Coating type				Excluded data lab:coating type
	P	IP	I	A	
General mean $m / \text{g.dm}^{-2}$	15,9	477,5	17,9	13,2	Lab 10:P Lab 8:I Lab 8:A
Repeatability std. dev. $s_r$	0,65	11,9	0,68	0,66	
Reproducibility std. dev. $s_R$	2,96	16,8	3,19	2,81	
Global repeatability std. dev.	$S_R = 0,0234 m + 0,6738 (R^2 \approx 1)$				
Global reproducibility std. dev.	$S_R = 0,0294 m + 2,7398 (R^2 \approx 1)$				
EN 12373-5 - Sealing quality assess. by admittance	Coating type				Excluded data lab:coating type
	P	IP	I	A	
General mean $m / Y$	8,98	n.a	15,29	9,88	Lab 12:P Lab 8:A
Repeatability std. dev. $s_r$	0,34	n.a	0,86	0,34	
Reproducibility std. dev. $s_R$	1,72	n.a	3,12	1,38	
Global repeatability std. dev.	0,51				
Global reproducibility std. dev.	$S_R = 0,1932 m (R^2=0,8218)$				

**Thickness**  
 Repeatability=0.56  
 Reproducibility=1.51  
 (standard :  
 1  $\mu\text{m}$  (until 10  $\mu\text{m}$  or 10%)

**Sealing/mass loss**  
 < 30  
 Repeatability=1  
 Reproducibility=3


**Sealing/admittance**  
 Repeatability=0.51  
 Reproducibility=3.12

# Estimation of loss of absorptive power of anodic oxidation coatings after sealing by dye spot test (EN 12373-4)



## Analysis of qualitative results

Frequency distribution of the results



Coating type	Dye absorption rating			Laboratory with results outside the range of 95%
	Mode	Median	Range of 95% of results	
P	0-1	0-1	0 to 2	None
IP	5	Between 4-5 and 5	3 to 5	Lab 2 (1 result)
A	0-1	0-1	0 to 1-2	None